

adinstbpm - Task #18055

Milestone # 17719 (New): Booster operational test of one crate alongside existing BPM system

No beam threshold

10/27/2017 08:51 AM - John Diamond

| | | | |
|--|--------------|------------------------|------------|
| Status: | Resolved | Start date: | 10/27/2017 |
| Priority: | Normal | Due date: | |
| Assignee: | John Diamond | % Done: | 100% |
| Category: | | Estimated time: | 2.00 hours |
| Target version: | | Spent time: | 1.00 hour |
| Description | | | |
| All of the other BPM systems have an intensity threshold that inhibits the position calculation. | | | |

History

#1 - 10/29/2017 09:58 AM - John Diamond

- Status changed from Assigned to Resolved
- % Done changed from 0 to 100

Implemented a threshold parameter in the BPM configuration section of bpmd.conf (the default is 50 counts). If the beam intensity is under the threshold we return NO_BEAM (999.99).